

INTEGRA METERING AG, Ringstrasse 75, CH-4106 Therwil

erklärt, dass das Produkt
declares that the product
déclare que le produit
dichiara che i prodotti

Durchflussgeber
Flowmeter
Débitmètre
Misuratore di portata

AMFLO® SONIC DryX III
ASDXIII

mit den Vorschriften folgender Richtlinien übereinstimmt :
conforms with the regulations of the following European Council Directives :
est conforme aux prescriptions et directives Européennes suivantes :
è conforme alle seguenti prescrizioni e direttive Europee :

CE Konformität/ CE compliance/ Conformité CE/ Conformità CE

Richtlinie Directive Direttiva		Beurteilungsverfahren Method of assessment Méthode d'évaluation Metodo di valutazione	Benannte Stelle Notified body Organisme notifié Organizzazione notificata
MID 2014/32/EU Messgeräte-Richtlinie Measurement Instruments Directive Directive sur les instruments de mesure Strumenti di misura direttiva		Module B: DE-25-MI004-PTB005 PTB, Berlin Normen/Standards: EN 1434:2022 OIML R75 :2002/2006	Module D: DE-M-AQ-PTB143 PTB – Cert 0102 Bundesallee 100 D - 38116 Braunschweig Module F: N/A
PED 2014/68/EU Druckgeräte-Richtlinie Pressure equipment directive Directive des équipements sous pression Direttiva di attrezzature a pressione		Regelwerk/relues/règles : AD 2000 01/2013	Module A2: Swiss Safety Center AG CE 1253 Richtstrasse 15 CH-8304 Wallisellen
2014/35/EU Niederspannungsrichtlinie Low voltage directive Directive basse tension Direttiva sulla bassa tensione		EN 61010-1 :2010 + A1 :2019 + A1 :2019/AC :2019 EN IEC 61010-2-030 :2021 + A11 :2021	
2014/30/EU EMV Richtlinie EMC Directive Directive CEM Direttiva EMC		EMC Normen / Standards: EN 61326-1 :2013 EN 1434-2022 Reports RCE-EMIESS24H268INT-01A RCE-EURESS24G953INT-01A	EMITECH Groupe 31300 TOULOUSE
2011/65/EU RoHS Richtlinie RoHS Directive Directive RoHS Direttiva RoHS		EN IEC 63000:2018	

Weitere Konformitäten

Richtlinie Directive Directive Direttiva	Beurteilungsverfahren Method of assessment Méthode d'évaluation Metodo di valutazione	Benannte Stelle Notified body Organisme notifié Organizzazione notificata
Zulassung national Deutschland: Mess- und Eichverordnung - MessEV vom 11.12.2014	TR K7.2 DIN EN 1434:2023 WELMEC 7.2 (2023) DE-25-M-PTB-0026 PTB, Berlin	Module D: DE-M-AQ-PTB143 PTB – Cert 0102 Bundesallee 100 D - 38116 Braunschweig

Unterzeichnet für und im Namen von:

Signed on behalf of:

Signé pour et au nom de :

Firmato per e al nome di:

Therwil, 05.12.2025

INTEGRA METERING AG



5 Dec 2025

Ilker AYCICEKLeiter Qualitätsmanagement
Head of Quality Management
Responsable Qualité
Responsabile Qualità**Christian Meier, INTEGRA Metering AG, CEO**
4106 Therwil, Switzerland, 5 Dec 2025

Simple electronic signature

Christian MEIERChief Executive Officer
Chief Executive Officer
Directeur général
Direttore generale